

# SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 2691  
CALIBRATION DATE: 28-Mar-03

SBE19 TEMPERATURE CALIBRATION DATA  
ITS-90 TEMPERATURE SCALE

## ITS-90 COEFFICIENTS

g = 4.26211268e-003  
h = 6.15957189e-004  
i = 9.88302967e-006  
j = -6.38064966e-007  
f0 = 1000.0

## ITS-68 COEFFICIENTS

a = 3.64763349e-003  
b = 5.94058978e-004  
c = 1.18523256e-005  
d = -6.37237185e-007  
f0 = 2759.960

BATH TEMP (ITS-90)	INSTRUMENT FREQ (Hz)	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
0.9999	2759.960	1.0001	0.00021
4.5000	2982.451	4.4996	-0.00041
15.0000	3726.424	15.0007	0.00069
18.5000	4000.923	18.4995	-0.00048
24.0000	4460.666	24.0000	-0.00002
29.0000	4909.468	28.9999	-0.00007
32.4999	5241.685	32.5000	0.00009

Temperature ITS-90 =  $1 / \{g + h[\ln(f_0/f)] + i[\ln^2(f_0/f)] + j[\ln^3(f_0/f)]\} - 273.15$  (°C)

Temperature ITS-68 =  $1 / \{a + b[\ln(f_0/f)] + c[\ln^2(f_0/f)] + d[\ln^3(f_0/f)]\} - 273.15$  (°C)

Following the recommendation of JPOTS:  $T_{68}$  is assumed to be  $1.00024 * T_{90}$  (-2 to 35 °C)

Residual = instrument temperature - bath temperature

